



Superoxide Dismutase SOD Assay Kit

Catalog # EA-7005

(For Research Use Only)

Introduction

The Superoxide Dismutase (SOD) Activity Assay utilizes WST-8 to assess SOD activity in biological samples. SOD is an enzyme that catalyzes the dismutation of the superoxide (O_2^-) anion radical into molecular oxygen (O_2) and hydrogen peroxide (H_2O_2). This assay indirectly measures the activity of SOD by using WST-8 to detect superoxide O_2^- levels in samples, which is suppressed when SOD activity is high. WST-8 interacts with superoxide O_2^- to form a colored product which can be measured spectrophotometrically at an absorbance of 450 nm.

Materials Required but Not Provided

- PBS
- 96-well clear microplate
- Microplate reader capable of measuring absorbance at 412 nm

Materials Provided

- WST Reagent (-80°C)
- Substrate Reagent (-20°C)
- 5mM DTPA (-20°C)
- Oxidase Reagent (4°C)

****Spin down small tubes before starting experiment. ****

Plasma Sample Preparation

1. Centrifuge citrated or EDTA-collected blood at 4°C ($1,000 \times g$ for 10 minutes) to separate plasma from erythrocytes. Alternatively, blood collected without anticoagulant can be centrifuged to collect serum.
2. Transfer the plasma layer to a new tube without disturbing the buffy layer.
3. The plasma may be assayed directly or stored away at -80°C .

Cell Sample Preparation

1. Wash the cells once with PBS before lysing the cells.
2. For a 96-well culture plate, add 40 μL of Lysis buffer to each well and incubate at room temperature for 10 minutes.
3. Pipette the Lysis buffer up and down to detach the cells and transfer the cell lysates into a new tube.
4. If necessary, homogenize the cell lysates with a sonicator.
5. The cell lysates may be assayed directly or stored at -80°C .

Tissue Sample Preparation

1. Weigh tissue sample and add 1 mL of Tissue Lysis buffer per 100mg of tissue.
2. Homogenize the tissue samples with a tissue grinder.
3. If necessary, further homogenize the tissue samples with a sonicator.
4. Centrifuge the sample at 10,000 RPM for 5 minutes to pellet the tissue debris.
5. Collect the supernatant and measure the protein concentration of the supernatant. The tissue sample can be assayed directly or stored at -80°C .
6. Use the Dilution buffer to dilute the tissue sample to the appropriate concentration for each assay.

SOD Measurement

1. Begin preparation of 10 mL of WST working solution by diluting 200 μ L WST reagent, 100 μ L Substrate reagent, and 100 μ L 10mM DTPA in 9.6 mL of PBS.
2. Right before loading the samples, complete the WST working solution by adding 10 μ L of the Oxidase reagent to the WST working solution and mixing. **Make sure the Oxidase reagent is evenly resuspended by pipetting up and down before using.**
3. In a clear 96-well plate, add 100 μ L of the WST working solution to each well. Add 10 μ L of cell sample to each well with WST working solution and mix thoroughly. **Be sure to load the samples quickly, since the WST reaction is active when the Oxidase reagent is added. Use a multichannel pipette if possible.**
4. For the control well, add 10 μ L of PBS to one of the wells with WST working solution.
5. Incubate the plate at 37°C for 45 minutes. The plate can be incubated for an additional hour or two if a stronger signal is desired.
6. Measure the absorbance of the plate at 450 nm using a plate reader.